

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3899

Name: RITTER 1  
Sterilizer ID: V2178198  
10/13/2023 09:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 179.2 F 0.0  
2:00 177.9 F 2.2  
4:00 175.5 F 5.9  
6:00 201.3 F 10.6  
8:00 216.9 F 4.7  
10:00 231.1 F 9.8  
12:00 236.9 F 8.8  
14:00 241.7 F 12.2  
16:00 240.4 F 10.6  
18:00 253.6 F 17.3  
20:00 265.1 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.6 F 29.2  
2:00 271.9 F 28.6  
3:00 271.3 F 28.5  
4:00 272.5 F 29.3

Min 271.2 F 28.2  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 215.7 F 0.0  
10:00 219.7 F 0.0  
15:00 220.1 F 0.0  
20:00 216.6 F 0.0  
25:00 216.6 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 21:50  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:00:42  
CYCLE COMPLETE

10/13/2023 10:28 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_