

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2306

Name: RITTER2  
Sterilizer ID: V2313228  
04/27/2023 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
2:00 192.3 F 3.5  
4:00 206.5 F 9.0  
6:00 223.6 F 8.7  
8:00 233.4 F 10.0  
10:00 235.6 F 8.9  
12:00 242.6 F 11.6  
14:00 247.3 F 14.0  
16:00 259.7 F 20.8  
18:00 270.8 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.4 F 29.2  
2:00 271.4 F 28.5  
3:00 272.4 F 29.4  
4:00 271.7 F 28.7  
  
Min 271.3 F 28.4  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 215.0 F 0.0  
10:00 216.1 F 0.0  
15:00 218.1 F 0.0  
20:00 211.9 F 0.0  
25:00 206.0 F 0.0  
30:00 199.7 F 0.0  
Drying Complete

FILLING: 2:24  
HEATING: 18:45  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:58:38  
CYCLE COMPLETE

04/27/2023 12:19 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_