
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006326
14:20 18/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:55
STERILIZING 2:52
136.4 C 223kPa 2:52
136.4 C 222kPa 3:22
136.4 C 224kPa 3:52
136.4 C 225kPa 4:22
136.4 C 224kPa 4:52
136.2 C 218kPa 5:22
136.4 C 224kPa 5:52
136.4 C 221kPa 6:22
136.4 C 221kPa 6:22
VENTING 6:23
AIR DRYING 7:15
CYCLE COMPLETE 8:50

Digital Signature #
4F428C42F5E56BEA
