
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001755
16:27 24/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:37
STERILIZING 3:32
136.0 C 220kPa 3:32
136.4 C 226kPa 4:02
136.2 C 223kPa 4:32
136.4 C 224kPa 5:02
136.3 C 224kPa 5:32
136.2 C 223kPa 6:02
136.4 C 224kPa 6:32
136.4 C 226kPa 7:02
136.4 C 226kPa 7:02
VENTING 7:02
AIR DRYING 7:54
CYCLE COMPLETE 25:56

Digital Signature #
0166B5901AA0C9F7
