

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002081  
10:52 23/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:41  
PRESSURIZING 2:43  
STERILIZING 3:47  
136.0 C 223kPa 3:47  
136.2 C 224kPa 4:17  
136.3 C 223kPa 4:47  
136.1 C 223kPa 5:17  
136.4 C 226kPa 5:47  
136.4 C 224kPa 6:17  
136.4 C 227kPa 6:47  
136.1 C 223kPa 7:17  
136.1 C 223kPa 7:17  
VENTING 7:17  
AIR DRYING 8:07  
CYCLE COMPLETE 8:18  
-----

Digital Signature #  
D8AA0CBB1CDEE237  
-----