
Midmark M11 - v1.0.5
Total Cycles: 2697

Name: RITTER2
Sterilizer ID: V2313228
07/25/2023 12:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.0 F 0.0
2:00 166.5 F 0.8
4:00 180.6 F 4.4
6:00 201.3 F 8.2
8:00 212.6 F 12.6
10:00 223.1 F 7.1
12:00 234.4 F 11.8
14:00 234.6 F 9.1
16:00 239.7 F 10.5
18:00 246.5 F 13.9
20:00 257.6 F 20.0
22:00 267.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.8 F 29.8
2:00 271.8 F 29.1
3:00 271.9 F 29.7
4:00 272.5 F 29.7

Min 271.3 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 209.7 F 0.0
10:00 207.7 F 0.0
15:00 205.0 F 0.0
20:00 201.7 F 0.0
25:00 198.3 F 0.0
30:00 196.1 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:02:25
CYCLE COMPLETE

07/25/2023 01:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____