
Midmark M11 - v1.0.0
Total Cycles: 4440

Name: RITTER 1
Sterilizer ID: V2178198
04/09/2024 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.5 F 0.0
2:00 98.5 F 0.8
4:00 110.0 F 2.9
6:00 141.3 F 5.9
8:00 165.3 F 9.5
10:00 205.1 F 9.1
12:00 208.9 F 5.4
14:00 220.2 F 9.0
16:00 231.3 F 13.2
18:00 233.6 F 9.2
20:00 239.8 F 10.7
22:00 243.9 F 12.5
24:00 255.7 F 18.6
26:00 266.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.5
3:00 272.5 F 29.2
4:00 271.6 F 28.5

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 200.0 F 0.0
10:00 200.8 F 0.0
15:00 197.1 F 0.0
20:00 194.4 F 0.0
25:00 198.2 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 27:37
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:06:10
CYCLE COMPLETE

04/09/2024 04:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____