
Midmark M11 - v1.0.5
Total Cycles: 194

Name: RITTER2
Sterilizer ID: V2313228
07/12/2021 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 160.6 F 2.6
4:00 166.1 F 6.2
6:00 190.2 F 10.6
8:00 210.8 F 5.6
10:00 218.4 F 10.1
12:00 230.6 F 8.5
14:00 237.6 F 12.5
16:00 239.7 F 11.0
18:00 251.4 F 17.0
20:00 262.1 F 23.2
22:00 272.0 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 271.9 F 30.2
2:00 271.8 F 30.4
3:00 272.1 F 30.4
4:00 272.1 F 30.2

Min 270.6 F 29.3
Max 273.2 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 215.5 F 0.0
10:00 221.2 F 0.0
15:00 223.7 F 0.0
20:00 225.9 F 0.0
25:00 227.3 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:29
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:58
CYCLE COMPLETE

07/12/2021 12:53 PM
ACCEPTED: X
REJECTED:
Operator: U01