
Midmark M11 - v1.0.5
Total Cycles: 1149

Name: RITTER2
Sterilizer ID: V2313228
06/01/2022 06:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.0 F 0.0
2:00 89.5 F 0.0
4:00 83.6 F 0.0
6:00 108.8 F 0.1
8:00 140.6 F 2.9
10:00 168.6 F 6.3
12:00 188.2 F 10.0
14:00 218.8 F 10.4
16:00 225.6 F 7.7
18:00 234.9 F 11.8
20:00 231.3 F 8.0
22:00 242.1 F 13.1
24:00 240.8 F 11.3
26:00 252.9 F 17.4
28:00 263.2 F 23.6
30:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.6 F 30.1
2:00 271.6 F 29.3
3:00 272.1 F 30.1
4:00 272.4 F 29.8

Min 271.2 F 29.2
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 234.0 F 0.0
10:00 227.1 F 0.0
15:00 219.3 F 0.0
20:00 198.9 F 0.0
25:00 198.9 F 0.0
30:00 197.5 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:06
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 01:08:57
CYCLE COMPLETE

06/01/2022 08:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____