
Midmark M11 - v1.0.0
Total Cycles: 1062

Name: RITTER 1
Sterilizer ID: V2178198
07/28/2021 06:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 74.8 F 0.0
4:00 108.2 F 0.0
6:00 121.2 F 0.1
8:00 148.9 F 0.1
10:00 172.2 F 0.2
12:00 194.4 F 0.4
14:00 208.7 F 3.6
16:00 217.3 F 6.0
18:00 226.0 F 8.9
20:00 235.1 F 12.4
22:00 232.9 F 8.1
24:00 243.9 F 13.0
26:00 241.3 F 10.9
28:00 237.7 F 9.2
30:00 251.0 F 15.5
32:00 261.7 F 21.6
34:00 272.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.7 F 28.6
2:00 271.5 F 27.9
3:00 272.7 F 28.7
4:00 271.8 F 28.0

Min 271.5 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 200.7 F 0.0
10:00 208.2 F 0.0
15:00 201.6 F 0.0
20:00 193.7 F 0.0
25:00 190.6 F 0.0
30:00 188.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 34:32
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:12:39
CYCLE COMPLETE

07/28/2021 08:14 AM
ACCEPTED: X
REJECTED:
Operator: U01