
Midmark M11 - v1.0.5
Total Cycles: 863

Name: RITTER2
Sterilizer ID: V2313228
03/09/2022 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 75.2 F 0.0
4:00 101.4 F 0.0
6:00 112.3 F 0.0
8:00 153.1 F 1.8
10:00 178.4 F 5.5
12:00 189.3 F 9.1
14:00 207.2 F 12.7
16:00 215.8 F 6.0
18:00 227.2 F 10.1
20:00 234.2 F 9.6
22:00 238.7 F 11.6
24:00 237.5 F 9.8
26:00 250.4 F 16.1
28:00 261.0 F 22.3
30:00 271.4 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.7 F 29.9
2:00 271.3 F 29.1
3:00 272.7 F 30.0
4:00 271.8 F 29.3

Min 271.3 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 208.5 F 0.0
10:00 203.9 F 0.0
15:00 190.0 F 0.0
20:00 184.1 F 0.0
25:00 184.1 F 0.0
30:00 186.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 30:41
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:08:44
CYCLE COMPLETE

03/09/2022 10:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____