
Midmark M11 - v1.0.0
Total Cycles: 4092

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2023 02:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.8 F 0.0
2:00 138.5 F 1.4
4:00 181.0 F 4.3
6:00 180.2 F 8.0
8:00 201.0 F 12.3
10:00 221.5 F 5.4
12:00 227.1 F 9.8
14:00 229.7 F 9.3
16:00 237.4 F 10.9
18:00 235.2 F 8.4
20:00 247.8 F 14.7
22:00 259.1 F 20.8
24:00 269.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.7 F 29.4
2:00 271.3 F 28.7
3:00 272.8 F 29.6
4:00 272.1 F 29.1

Min 271.2 F 28.7
Max 273.4 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 239.2 F 0.0
10:00 232.5 F 0.0
15:00 226.8 F 0.0
20:00 231.5 F 0.0
25:00 227.7 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:02
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:49
CYCLE COMPLETE

12/06/2023 04:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____