
Midmark M11 - v1.0.5
Total Cycles: 1603

Name: RITTER2
Sterilizer ID: V2313228
10/17/2022 07:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 76.9 F 0.0
4:00 103.8 F 0.0
6:00 124.3 F 0.1
8:00 144.5 F 0.2
10:00 173.8 F 1.0
12:00 193.5 F 4.1
14:00 209.3 F 7.6
16:00 221.3 F 11.1
18:00 225.9 F 6.7
20:00 235.2 F 10.6
22:00 236.4 F 8.5
24:00 243.9 F 13.3
26:00 242.6 F 12.0
28:00 254.4 F 18.1
30:00 264.9 F 24.5
32:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.7
2:00 271.8 F 29.0
3:00 271.7 F 29.4
4:00 272.4 F 29.5

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 202.2 F 0.0
10:00 209.3 F 0.0
15:00 211.1 F 0.0
20:00 213.9 F 0.0
25:00 214.1 F 0.0
30:00 215.5 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 32:00
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

10/17/2022 08:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____