
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000790
8:09 30/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:20
STERILIZING 3:23
136.0 C 227kPa 3:23
136.4 C 231kPa 3:53
136.1 C 229kPa 4:23
136.4 C 229kPa 4:53
136.4 C 233kPa 5:23
136.1 C 229kPa 5:53
136.3 C 231kPa 6:23
136.4 C 232kPa 6:53
136.4 C 232kPa 6:53
VENTING 6:53
AIR DRYING 7:44
CYCLE COMPLETE 7:46

Digital Signature #
0AA85A9217FC2C92
