
Midmark M11 - v1.0.0
Total Cycles: 1472

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2021 08:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 89.3 F 0.0
4:00 80.3 F 0.0
6:00 105.2 F 0.1
8:00 137.5 F 0.8
10:00 167.9 F 2.4
12:00 191.4 F 4.4
14:00 206.7 F 7.1
16:00 217.7 F 10.3
18:00 227.5 F 8.8
20:00 230.9 F 8.0
22:00 241.9 F 12.5
24:00 238.5 F 9.7
26:00 238.9 F 9.2
28:00 248.9 F 14.3
30:00 260.3 F 20.6
32:00 270.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.7 F 28.8
2:00 272.4 F 28.3
3:00 271.6 F 27.8
4:00 271.8 F 28.3

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 220.0 F 0.0
10:00 194.4 F 0.0
15:00 196.8 F 0.0
20:00 191.6 F 0.0
25:00 194.4 F 0.0
30:00 203.4 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 32:51
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:10:55
CYCLE COMPLETE

12/20/2021 09:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____