
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001884
8:57 24/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:46
STERILIZING 3:58
136.4 C 224kPa 3:58
136.3 C 225kPa 4:28
136.4 C 225kPa 4:58
136.4 C 225kPa 5:28
136.4 C 222kPa 5:58
136.4 C 225kPa 6:28
136.1 C 221kPa 6:58
136.4 C 224kPa 7:28
136.4 C 224kPa 7:28
VENTING 7:28
AIR DRYING 8:24
CYCLE COMPLETE 12:12

Digital Signature #
7B09FBBA52ACE515
