
Midmark M11 - v1.0.0
Total Cycles: 4934

Name: RITTER 1
Sterilizer ID: V2178198
09/24/2024 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.6 F 0.0
2:00 110.4 F 0.0
4:00 121.5 F 0.0
6:00 152.2 F 0.1
8:00 185.1 F 0.3
10:00 201.9 F 0.9
12:00 214.9 F 4.2
14:00 222.0 F 7.4
16:00 230.9 F 11.0
18:00 232.2 F 7.3
20:00 241.4 F 11.9
22:00 238.4 F 9.8
24:00 238.6 F 9.2
26:00 249.4 F 14.7
28:00 260.4 F 20.8
30:00 270.3 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 271.7 F 28.1
2:00 272.3 F 28.6
3:00 272.8 F 28.8
4:00 272.6 F 28.6

Min 271.6 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 193.9 F 0.0
10:00 192.1 F 0.0
15:00 188.1 F 0.0
20:00 217.1 F 0.0
25:00 224.2 F 0.0
30:00 235.8 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:54
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:09:26
CYCLE COMPLETE

09/24/2024 11:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____