
Midmark M11 - v1.0.5
Total Cycles: 515

Name: RITTER2
Sterilizer ID: V2313228
11/03/2021 06:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 74.2 F 0.0
4:00 77.8 F 0.0
6:00 98.2 F 0.1
8:00 148.9 F 0.2
10:00 186.5 F 0.4
12:00 207.6 F 3.8
14:00 218.3 F 7.3
16:00 237.2 F 11.1
18:00 230.4 F 7.5
20:00 244.9 F 13.0
22:00 242.3 F 11.6
24:00 240.7 F 10.7
26:00 254.5 F 17.7
28:00 266.1 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.4 F 29.3
2:00 271.5 F 28.4
3:00 272.4 F 29.3
4:00 271.6 F 28.5

Min 271.1 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 234.8 F 0.0
10:00 231.7 F 0.0
15:00 228.1 F 0.0
20:00 239.1 F 0.0
25:00 228.4 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:28
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:07:16
CYCLE COMPLETE

11/03/2021 07:50 AM
ACCEPTED: X
REJECTED:
Operator: U01