

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1664

Name: RITTER 1  
Sterilizer ID: V2178198  
02/23/2022 04:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.2 F 0.0  
2:00 157.6 F 2.3  
4:00 178.9 F 6.0  
6:00 199.5 F 10.7  
8:00 215.9 F 4.5  
10:00 229.0 F 9.5  
12:00 234.9 F 9.1  
14:00 241.9 F 11.9  
16:00 240.0 F 10.3  
18:00 253.3 F 17.0  
20:00 264.9 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.7 F 29.0  
2:00 271.9 F 28.3  
3:00 271.8 F 28.5  
4:00 272.8 F 28.9  
  
Min 271.4 F 28.0  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.4 F 0.0  
5:00 224.3 F 0.0  
10:00 231.1 F 0.0  
15:00 227.4 F 0.0  
20:00 223.0 F 0.0  
25:00 222.7 F 0.0  
30:00 222.6 F 0.0  
Drying Complete

FILLING: 1:29  
HEATING: 21:52  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:29  
CYCLE COMPLETE

02/23/2022 05:31 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_