
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005454
13:57 01/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:15
PRESSURIZING 2:16
STERILIZING 3:16
136.0 C 220kPa 3:16
136.4 C 225kPa 3:46
136.3 C 225kPa 4:16
136.4 C 225kPa 4:46
136.0 C 220kPa 5:16
136.4 C 220kPa 5:46
136.4 C 223kPa 6:16
136.4 C 224kPa 6:46
136.4 C 224kPa 6:46
VENTING 6:46
AIR DRYING 7:39
CYCLE COMPLETE 8:15

Digital Signature #
A8025B31D4A92621
