

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5282

Name: RITTER 1  
Sterilizer ID: V2178198  
01/27/2025 05:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 96.4 F 0.0  
2:00 113.0 F 0.0  
4:00 133.9 F 1.9  
6:00 147.3 F 4.7  
8:00 170.3 F 7.9  
10:00 186.7 F 11.3  
12:00 195.5 F 4.3  
14:00 211.7 F 6.6  
16:00 222.6 F 10.2  
18:00 230.4 F 9.7  
20:00 234.7 F 10.6  
22:00 234.5 F 8.4  
24:00 244.6 F 13.6  
26:00 255.5 F 19.2  
28:00 265.6 F 25.3  
30:00 271.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 272.1 F 29.8  
2:00 272.9 F 30.0  
3:00 272.4 F 29.7  
4:00 272.0 F 29.5

Min 271.4 F 29.2  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
5:00 200.2 F 0.0  
10:00 195.7 F 0.0  
15:00 186.4 F 0.0  
20:00 181.6 F 0.0  
25:00 182.2 F 0.0  
30:00 194.3 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:02  
STERILIZING: 4:00  
VENTING: 4:01  
DRYING: 30:00  
TOTAL CYCLE: 01:08:52  
CYCLE COMPLETE

01/27/2025 06:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_