
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 005068
12:05 06/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:04
STERILIZING 4:16
136.4 C 222kPa 4:16
136.4 C 224kPa 4:46
136.4 C 224kPa 5:16
136.4 C 223kPa 5:46
136.2 C 222kPa 6:16
136.3 C 222kPa 6:46
136.4 C 224kPa 7:16
136.3 C 223kPa 7:46
136.3 C 223kPa 7:46
VENTING 7:46
AIR DRYING 8:40
CYCLE COMPLETE 14:25

Digital Signature #
B52C8F687BB23948
