
Midmark M11 - v1.0.5
Total Cycles: 3910

Name: RITTER2
Sterilizer ID: V2313228
09/06/2024 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 69.0 F 0.0
4:00 77.7 F 0.0
6:00 101.9 F 0.1
8:00 151.3 F 0.4
10:00 177.3 F 2.6
12:00 195.3 F 6.1
14:00 211.1 F 9.7
16:00 224.0 F 13.5
18:00 226.3 F 7.8
20:00 237.0 F 12.2
22:00 236.1 F 9.5
24:00 239.0 F 9.8
26:00 247.3 F 14.2
28:00 258.9 F 20.6
30:00 269.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 29.1
2:00 272.2 F 29.5
3:00 271.7 F 28.6
4:00 272.5 F 29.4

Min 271.5 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 227.5 F 0.0
10:00 203.5 F 0.0
15:00 195.9 F 0.0
20:00 192.4 F 0.0
25:00 204.7 F 0.0
30:00 205.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:02
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:07
CYCLE COMPLETE

09/06/2024 10:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____