

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2890

Name: RITTER 1  
Sterilizer ID: V2178198  
02/13/2023 11:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.8 F 0.0  
2:00 195.6 F 3.7  
4:00 195.1 F 8.3  
6:00 214.0 F 11.4  
8:00 226.6 F 8.6  
10:00 237.5 F 11.2  
12:00 240.4 F 11.6  
14:00 238.8 F 10.1  
16:00 252.5 F 16.9  
18:00 264.3 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.6 F 29.3  
2:00 271.6 F 28.6  
3:00 272.7 F 29.4  
4:00 272.0 F 28.8  
  
Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 233.3 F 0.0  
10:00 226.9 F 0.0  
15:00 230.5 F 0.0  
20:00 228.6 F 0.0  
25:00 231.2 F 0.0  
30:00 224.3 F 0.0  
  
Drying Complete

FILLING: 1:47  
HEATING: 19:58  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

02/13/2023 12:05 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_