
Midmark M11 - v1.0.5
Total Cycles: 676

Name: RITTER2
Sterilizer ID: V2313228
01/10/2022 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 171.1 F 3.1
4:00 179.2 F 7.1
6:00 203.2 F 12.5
8:00 220.8 F 7.2
10:00 232.9 F 12.4
12:00 236.1 F 10.1
14:00 236.2 F 8.9
16:00 249.2 F 15.4
18:00 261.2 F 22.2
20:00 272.2 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.5
2:00 271.4 F 29.1
3:00 272.5 F 29.6
4:00 271.5 F 28.9

Min 271.3 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 212.0 F 0.0
10:00 222.2 F 0.0
15:00 224.6 F 0.0
20:00 235.7 F 0.0
25:00 230.7 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:54
HEATING: 20:31
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

01/10/2022 01:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____