
Midmark M11 - v1.0.0
Total Cycles: 2415

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2022 10:57 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.9 F 0.0
2:00 166.9 F 3.6
4:00 186.2 F 7.9
6:00 202.9 F 9.0
8:00 224.3 F 8.2
10:00 235.0 F 9.9
12:00 241.9 F 11.9
14:00 240.5 F 10.4
16:00 255.0 F 17.7
18:00 267.4 F 25.3

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 217F
CHAMBER PRESS 1.4PSI
** ITEMS NOT STERILE **

10/11/2022 11:16 AM