
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002199
13:34 01/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:02
STERILIZING 3:02
136.4 C 224kPa 3:02
135.8 C 222kPa 3:32
136.0 C 222kPa 4:02
136.3 C 225kPa 4:32
136.4 C 226kPa 5:02
136.3 C 223kPa 5:32
136.4 C 225kPa 6:02
136.4 C 226kPa 6:32
136.4 C 226kPa 6:32
VENTING 6:32
AIR DRYING 7:22
CYCLE COMPLETE 17:40

Digital Signature #
28EC4218396F071F
