

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001441  
10:29 07/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:40  
STERILIZING 3:47  
136.5 C 225kPa 3:47  
136.4 C 226kPa 4:17  
136.2 C 224kPa 4:47  
136.4 C 225kPa 5:17  
136.2 C 222kPa 5:47  
136.1 C 223kPa 6:17  
136.4 C 222kPa 6:47  
136.4 C 226kPa 7:17  
136.4 C 226kPa 7:17  
VENTING 7:17  
AIR DRYING 8:07  
CYCLE COMPLETE 8:12

-----  
Digital Signature #  
C6F8EB5261C1F69E  
-----