
Midmark M11 - v1.0.0
Total Cycles: 5252

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2025 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 75.3 F 0.0
4:00 108.4 F 0.0
6:00 119.0 F 0.0
8:00 145.0 F 0.2
10:00 174.0 F 0.3
12:00 194.8 F 0.7
14:00 212.6 F 2.9
16:00 221.6 F 5.9
18:00 229.8 F 9.1
20:00 239.2 F 12.9
22:00 236.9 F 9.5
24:00 239.1 F 9.6
26:00 246.4 F 13.3
28:00 243.5 F 11.8
30:00 255.9 F 18.2
32:00 267.1 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.7 F 28.6
2:00 271.5 F 27.8
3:00 272.8 F 28.7
4:00 271.9 F 28.0

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.7 F 0.0
5:00 199.2 F 0.0
10:00 196.7 F 0.0
15:00 222.4 F 0.0
20:00 228.3 F 0.0
25:00 226.1 F 0.0
30:00 229.4 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 33:27
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:11:37
CYCLE COMPLETE

01/17/2025 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____