
Midmark M11 - v1.0.0
Total Cycles: 5161

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2024 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 73.9 F 0.0
4:00 89.0 F 0.0
6:00 121.6 F 0.0
8:00 149.5 F 0.2
10:00 174.0 F 0.5
12:00 194.3 F 0.9
14:00 211.0 F 3.8
16:00 220.5 F 7.0
18:00 228.8 F 10.5
20:00 234.6 F 8.7
22:00 239.9 F 11.2
24:00 235.7 F 8.8
26:00 239.7 F 10.0
28:00 247.2 F 13.9
30:00 258.6 F 20.0
32:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.3 F 29.0
2:00 272.8 F 29.1
3:00 272.2 F 28.6
4:00 271.6 F 28.3

Min 271.6 F 28.3
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 221.1 F 0.0
10:00 230.6 F 0.0
15:00 226.4 F 0.0
20:00 233.9 F 0.0
25:00 227.0 F 0.0
30:00 228.0 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 33:17
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:11:32
CYCLE COMPLETE

12/05/2024 10:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____