

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005943  
9:32 11/07/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:45  
PRESSURIZING 1:52  
STERILIZING 2:56  
136.4 C 224kPa 2:56  
136.4 C 223kPa 3:26  
136.4 C 224kPa 3:56  
136.4 C 224kPa 4:26  
136.0 C 222kPa 4:56  
136.3 C 224kPa 5:26  
136.4 C 224kPa 5:56  
136.3 C 223kPa 6:26  
136.3 C 223kPa 6:26  
VENTING 6:26  
AIR DRYING 7:16  
CYCLE COMPLETE 7:18  
-----

Digital Signature #  
873701D2CFEE8DEE  
-----