

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000275  
12:37 01/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:44  
PRESSURIZING 3:46  
STERILIZING 5:19  
136.0 C 229kPa 5:19  
136.2 C 229kPa 5:49  
136.3 C 233kPa 6:19  
136.3 C 233kPa 6:49  
136.3 C 232kPa 7:19  
136.4 C 233kPa 7:49  
136.2 C 232kPa 8:19  
136.0 C 231kPa 8:49  
136.0 C 231kPa 8:49  
VENTING 8:50  
AIR DRYING 9:51  
CYCLE COMPLETE 24:51  
-----

Digital Signature #  
F56C5AA55527B19C  
-----