

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2031

Name: RITTER 1  
Sterilizer ID: V2178198  
06/16/2022 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
2:00 182.4 F 3.8  
4:00 196.7 F 8.4  
6:00 223.6 F 10.9  
8:00 228.0 F 8.9  
10:00 237.6 F 10.3  
12:00 241.8 F 11.7  
14:00 238.9 F 10.0  
16:00 252.7 F 16.7  
18:00 264.0 F 23.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 271.7 F 28.5  
2:00 272.3 F 29.0  
3:00 271.3 F 28.3  
4:00 272.0 F 28.7  
  
Min 271.1 F 28.2  
Max 272.4 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.8 F 0.0  
5:00 229.4 F 0.0  
10:00 233.8 F 0.0  
15:00 220.5 F 0.0  
20:00 215.1 F 0.0  
25:00 223.8 F 0.0  
30:00 230.4 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 19:58  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:58  
CYCLE COMPLETE

06/16/2022 01:53 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_