
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000384
13:49 22/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:50
STERILIZING 3:58
136.0 C 228kPa 3:58
136.4 C 230kPa 4:28
136.1 C 230kPa 4:58
136.4 C 232kPa 5:28
136.3 C 232kPa 5:58
136.4 C 233kPa 6:28
136.3 C 231kPa 6:58
136.3 C 232kPa 7:28
136.3 C 232kPa 7:28
VENTING 7:28
AIR DRYING 8:20
CYCLE COMPLETE 16:11

Digital Signature #
C5306DE0A6EBEDE3
