
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 000310
10:47 20/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:18
STERILIZING 4:31
136.0 C 228kPa 4:31
136.0 C 230kPa 5:01
136.3 C 232kPa 5:31
136.3 C 233kPa 6:01
136.3 C 233kPa 6:31
136.4 C 233kPa 7:01
136.4 C 233kPa 7:31
136.3 C 232kPa 8:01
136.3 C 232kPa 8:01
VENTING 8:01
AIR DRYING 8:54
CYCLE COMPLETE 23:54

Digital Signature #
EB4DEF2B68664C26
