

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001780  
9:35 16/05/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:32  
PRESSURIZING 2:38  
STERILIZING 3:43  
136.0 C 220kPa 3:43  
136.4 C 224kPa 4:13  
136.1 C 221kPa 4:43  
136.2 C 222kPa 5:13  
136.3 C 220kPa 5:43  
136.3 C 222kPa 6:13  
136.4 C 225kPa 6:43  
136.4 C 224kPa 7:13  
136.4 C 224kPa 7:13  
VENTING 7:13  
AIR DRYING 8:03  
CYCLE COMPLETE 12:13  
-----

Digital Signature #  
CC7D942CEF9C3C22  
-----