

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001576  
14:15 11/11/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:55  
PRESSURIZING 6:28  
STERILIZING 8:01  
136.0 C 220kPa 8:01  
136.4 C 218kPa 8:31  
136.3 C 221kPa 9:01  
136.2 C 221kPa 9:31  
136.5 C 222kPa 10:01  
136.4 C 222kPa 10:31  
136.3 C 222kPa 11:01  
136.4 C 222kPa 11:31  
VENTING 11:31  
AIR DRYING 12:31  
CYCLE COMPLETE 20:51

-----  
Digital Signature #  
DA80926847604925  
-----