
Midmark M11 - v1.0.5
Total Cycles: 2036

Name: RITTER2
Sterilizer ID: V2313228
02/09/2023 06:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.1 F 0.0
2:00 197.5 F 3.9
4:00 214.2 F 9.8
6:00 225.5 F 7.0
8:00 238.7 F 12.8
10:00 240.9 F 11.8
12:00 242.0 F 11.7
14:00 256.8 F 19.4
16:00 269.2 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.3 F 29.3
2:00 272.2 F 29.7
3:00 272.1 F 29.1
4:00 272.3 F 29.7

Min 271.5 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 222.9 F 0.0
10:00 228.6 F 0.0
15:00 227.4 F 0.0
20:00 229.5 F 0.0
25:00 227.0 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 2:49
HEATING: 16:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:01
CYCLE COMPLETE

02/09/2023 07:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____