

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007493  
13:00 04/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:03  
STERILIZING 4:08  
136.4 C 222kPa 4:08  
136.1 C 222kPa 4:38  
136.4 C 222kPa 5:08  
136.4 C 220kPa 5:38  
136.4 C 225kPa 6:08  
136.4 C 223kPa 6:38  
136.4 C 223kPa 7:08  
136.4 C 224kPa 7:38  
136.4 C 224kPa 7:38  
VENTING 7:38  
AIR DRYING 8:33  
CYCLE COMPLETE 10:56

-----  
Digital Signature #  
6FC6F7AEBC7B38C6  
-----