
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000484
15:12 19/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:48
PRESSURIZING 3:50
STERILIZING 4:58
136.3 C 227kPa 4:58
136.4 C 232kPa 5:28
136.3 C 231kPa 5:58
136.3 C 232kPa 6:28
136.4 C 231kPa 6:58
136.3 C 232kPa 7:28
136.3 C 232kPa 7:58
136.4 C 233kPa 8:28
136.4 C 233kPa 8:28
VENTING 8:28
CYCLE ABORTED 8:48
STOP BUTTON PRESSED
NOT STERILE
END TIME 8:48

Digital Signature #
D3A2E89C1B997DB2
