

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1028

Name: RITTER2  
Sterilizer ID: V2313228  
04/25/2022 02:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.5 F 0.0  
2:00 171.5 F 3.2  
4:00 180.7 F 7.4  
6:00 206.5 F 12.4  
8:00 219.9 F 7.0  
10:00 235.2 F 12.2  
12:00 235.0 F 9.7  
14:00 238.6 F 9.2  
16:00 248.6 F 15.2  
18:00 260.8 F 22.1  
20:00 272.0 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.4 F 29.5  
2:00 271.8 F 29.5  
3:00 272.3 F 29.4  
4:00 271.6 F 29.4

Min 271.5 F 29.0  
Max 272.5 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 230.3 F 0.0  
10:00 225.2 F 0.0  
15:00 220.5 F 0.0  
20:00 210.8 F 0.0  
25:00 207.8 F 0.0  
30:00 205.2 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 20:30  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

04/25/2022 03:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_