
Midmark M11 - v1.0.5
Total Cycles: 2817

Name: RITTER2
Sterilizer ID: V2313228
08/30/2023 10:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 164.9 F 0.1
4:00 186.5 F 3.4
6:00 204.1 F 7.5
8:00 219.6 F 12.4
10:00 230.0 F 8.7
12:00 238.8 F 11.2
14:00 241.8 F 12.3
16:00 241.5 F 11.4
18:00 254.2 F 17.9
20:00 265.3 F 24.6
22:00 271.4 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.5 F 29.7
2:00 272.8 F 30.1
3:00 271.6 F 29.4
4:00 273.0 F 29.9

Min 271.5 F 29.4
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 230.1 F 0.0
10:00 195.6 F 0.0
15:00 193.2 F 0.0
20:00 189.4 F 0.0
25:00 185.9 F 0.0
30:00 183.2 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 22:34
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

08/30/2023 11:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____