
Midmark M11 - v1.0.5
Total Cycles: 1662

Name: RITTER2
Sterilizer ID: V2313228
10/26/2022 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.0 F 0.0
2:00 187.3 F 3.9
4:00 212.2 F 9.1
6:00 226.1 F 6.8
8:00 232.3 F 10.2
10:00 235.1 F 8.2
12:00 241.2 F 11.3
14:00 245.8 F 13.8
16:00 258.4 F 20.7
18:00 269.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.3 F 29.7
2:00 271.7 F 29.7
3:00 272.2 F 29.8
4:00 271.5 F 29.4

Min 271.4 F 29.2
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 224.5 F 0.0
10:00 237.6 F 0.0
15:00 232.7 F 0.0
20:00 228.5 F 0.0
25:00 225.9 F 0.0
30:00 225.0 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 18:57
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:02
CYCLE COMPLETE

10/26/2022 02:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____