

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.9uS / 0.5ppm  
CYCLE NUMBER 001078  
8:03 17/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:35  
STERILIZING 3:40  
136.0 C 220kPa 3:40  
136.2 C 224kPa 4:10  
136.4 C 224kPa 4:40  
136.4 C 224kPa 5:10  
136.4 C 225kPa 5:40  
136.3 C 224kPa 6:10  
136.3 C 223kPa 6:40  
136.3 C 224kPa 7:10  
136.3 C 224kPa 7:10  
VENTING 7:10  
AIR DRYING 8:01  
CYCLE COMPLETE 23:01  
-----

Digital Signature #  
577C245F368D3F57  
-----