

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000471  
9:44 12/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:47  
PRESSURIZING 3:49  
STERILIZING 4:57  
136.0 C 228kPa 4:57  
136.1 C 230kPa 5:27  
136.3 C 231kPa 5:57  
136.4 C 230kPa 6:27  
136.4 C 233kPa 6:57  
136.3 C 230kPa 7:27  
136.4 C 232kPa 7:57  
136.4 C 232kPa 8:27  
136.4 C 232kPa 8:27  
VENTING 8:27  
AIR DRYING 9:21  
CYCLE COMPLETE 23:40

-----  
Digital Signature #  
BAB73D05DE3A47EF  
-----