

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004416  
8:25 22/09/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:03  
PRESSURIZING 3:05  
STERILIZING 4:04  
136.0 C 220kPa 4:04  
136.4 C 224kPa 4:34  
136.2 C 222kPa 5:04  
136.2 C 222kPa 5:34  
135.9 C 219kPa 6:04  
136.3 C 223kPa 6:34  
136.4 C 223kPa 7:04  
136.4 C 223kPa 7:34  
136.4 C 223kPa 7:34  
VENTING 7:34  
AIR DRYING 8:19  
CYCLE COMPLETE 11:02

-----

Digital Signature #  
63C5931F0360359C  
-----