
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001410
15:57 25/03/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:53
STERILIZING 5:04
136.0 C 220kPa 5:04
136.3 C 224kPa 5:35
136.3 C 221kPa 6:05
136.3 C 223kPa 6:35
136.3 C 223kPa 7:05
136.1 C 223kPa 7:35
136.4 C 224kPa 8:05
136.4 C 226kPa 8:35
136.4 C 226kPa 8:35
VENTING 8:35
AIR DRYING 9:38
CYCLE COMPLETE 15:10

Digital Signature #
D895237ABBC211A4
