
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000452
11:27 01/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:05
STERILIZING 4:14
136.0 C 228kPa 4:14
136.2 C 232kPa 4:44
136.3 C 233kPa 5:14
136.4 C 233kPa 5:44
136.4 C 232kPa 6:14
136.2 C 232kPa 6:44
136.4 C 233kPa 7:14
136.3 C 233kPa 7:44
136.3 C 233kPa 7:44
VENTING 7:44
AIR DRYING 8:40
CYCLE COMPLETE 8:40

Digital Signature #
30A7485717DFB550
