
Midmark M11 - v1.0.0
Total Cycles: 2713

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2022 04:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.0 F 0.0
2:00 200.1 F 4.1
4:00 207.2 F 8.9
6:00 221.7 F 9.0
8:00 228.7 F 9.3
10:00 236.0 F 9.4
12:00 242.6 F 12.4
14:00 241.2 F 10.9
16:00 254.4 F 17.6
18:00 265.7 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.6 F 28.9
2:00 272.1 F 28.9
3:00 272.3 F 28.7
4:00 272.1 F 28.9

Min 271.7 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 221.5 F 0.0
10:00 232.7 F 0.0
15:00 228.3 F 0.0
20:00 225.4 F 0.0
25:00 224.3 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 19:41
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

12/21/2022 05:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____