

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4190

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2024 03:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 111.8 F 0.0  
2:00 114.2 F 1.1  
4:00 161.4 F 3.6  
6:00 168.5 F 6.9  
8:00 188.2 F 10.6  
10:00 213.5 F 4.9  
12:00 212.6 F 6.9  
14:00 225.6 F 11.1  
16:00 228.1 F 6.9  
18:00 238.9 F 12.4  
20:00 238.9 F 10.6  
22:00 251.0 F 16.5  
24:00 261.3 F 22.6  
26:00 271.4 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.4 F 29.8  
2:00 271.8 F 29.6  
3:00 272.5 F 29.7  
4:00 271.6 F 29.5

Min 271.4 F 29.2  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.8 F 0.0  
5:00 222.5 F 0.0  
10:00 225.4 F 0.0  
15:00 212.6 F 0.0  
20:00 208.3 F 0.0  
25:00 214.4 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 26:38  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:05:26  
CYCLE COMPLETE

01/10/2024 04:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_